

Instructions for use of Isohelix SK2 Buccal Swabs

(REF SK-2S)

Intended for the retrieval of buccal cells. Single use only.

Store at room temperature. Use only if swab wrapper remains intact.



Follow the correct sampling procedure to avoid risk of the swab head becoming detached in the mouth. If the swab head does become detached from the swab stick whilst in the mouth, remove immediately.

<p>Take your DNA sample at least 1 hour after eating, drinking or cleaning your teeth. For best results, rinse mouth with water immediately prior to sampling.</p>	
<p>1</p>	<p>Pull open the package from one end.</p>
<p>2</p>	<p>Remove the swab from the tube, taking care not to touch the white swab head with your fingers</p>
<p>3</p>	<p>Insert the swab into your mouth and rub firmly against the inside of your cheek or underneath lower and upper lip. For standard DNA collection rub for 1 minute and in all cases rub for a minimum of 20 seconds. Important – use reasonable, firm and solid pressure</p>
<p>4</p>	<p>Slide the plastic cap over the swab handle with the flat side of the cap facing upwards and the swab facing downwards.</p>
<p>5</p>	<p>Insert the swab into the clear plastic tube and push the cap into place. Next, hold the cap while pulling the swab handle outwards to release the swab material into the tube.</p>
<p>6</p>	<p>Close the cap by pushing the stopper fully into the cap ensuring the stopper is flush with the cap. The tube is now completely sealed.</p>

You may now proceed directly to isolate the DNA from the swab sample. If the swab needs to be stored prior to DNA isolation, steps should be taken to stabilise the DNA on the swab to maximise yield and quality. Advice and options for DNA stabilisation can be found at www.isohelix.com/support.html

For Research Use only

Isohelix is a division of Cell Projects

For Swab or DNA Isolation queries info@isohelix.com www.isohelix.com

Molecular Biology Solutions www.cellprojects.com

Unit 2 Roebuck Business Park, Harrietsham, Kent, ME17 1AB, UK ☎: +44 1622 851177